

# Temperature Measurement of Various Permittivity Medium by AC ECT

K.Manikandan<sup>1</sup>, S.Sathiyamoorthy<sup>2</sup>

<sup>12</sup>JJ College of Engineering and Technology, Tiruchirappalli, India.

## Abstract

The measurement of temperature is important in various sectors. In this proposed method, AC Electrical Capacitance Tomography (ECT) is used to detect the temperature of the container with contact less of the medium. In our system, depends on permittivity distribution of a dielectric medium inside the tank. The ECT electrodes are mounted outside the tank and it is act as sensor. The dielectric permittivity changes, which temperature rises of the medium in the container. In according to this method the ECT sensor output varies. Ac signal is applied to the ECT. Based on the dielectric medium inside the tank, the electrodes exhibit equivalent current signal of the medium and signal conditioning circuits are used to convert electrode current in to voltage. Finally, the liquid and solid medium temperature can be measured from the AC ECT system. It is a non-contact, non-intrusive, radiation free, simple and low cost technique.

**Keywords:** *Electrical Capacitance Tomography, temperature measurement, permittivity, dielectric medium.*

## 1. INTRODUCTION

The Electrical Capacitance Tomography generally used for flow measurement. In this projected work temperature measures with the help of ECT principle. The medium range temperature is sensed and equivalent voltages are exhibits to each set of electrodes with necessary circuits. In process industries temperature sensor and flow sensor widely used. The proposed method tells that single sensor is sensing both parameters at a time after calibration. The experimental study analyzing

the output voltages and input temperature, based on changes of

permittivity medium. ECT is a technique for measuring the permittivity distribution of a mixture of two dielectric materials inside a closed vessel, from which the concentration distribution can be found. They have application in industry, medical imaging, geophysics, biology etc. Industrial Tomography is an emerging technique used to visualize internal behavior of industrial processes such as temperature, level in tank, gas/ oil/ water flows in pipelines, pneumatic conveyors and in mixing/separation processes (1).

The temperature measurement in industry is a complex and random mechanism Modulation of light signal takes place in side fibres sensor based on change. Measurement is done through Fibres Electro Optic active sensor the biological sensor (cell) constructed around the fibres optic probe in which 3ml of water solution and saline it held the cell constant. Optical signal absorption rate is varied based on change in biological cell then Electro Optic technique is used to convert light signal to electrical is proportional to change in (2,3).The sensor is made-upon UV molding process with simple and safe procedures. The overall sensor size is around 150  $\mu\text{m}$  in diameter and 343  $\mu\text{m}$  in length. It is used in biomedical applications to simple and pressure. The change cavity length and diameter sensed by change in intensity of wave length of UV signal. The range signal measurement is less and very sensitive type. (7). The Brine activity of human being is done through capacitance measurement the permittivity of capacitance is varied by varied

change of in head. The 21 electrode is placed on head to measure the activity brine .Electrical impedance method is used earlier bio medical application in which sensitivity of measurement is respect to voltage to current ratio then the operating range is minimum because change of in brine is very less. In this method they use Electrical Capacitance Tomography to analysis the brine activity image reconstruction was done various algorithms [10]. The MEMS technology is used to analyze the in process industry in which cascade triple bend the beam structure. The sensitivity of the sensor is increased through cascade. The sensor contains LC circuit in which resonant frequency of sensor varied based on field. The thermal energy converted into mechanical energy then the mechanical energy converted to electrical energy of resonant circuit. Triple beam cascade structure contain electrode parallel plates that are embedded on substrate and movable beam whenever the change occur in the field then the plate get deformed due to the deformation the capacitor value changed. The change in capacitor value is proportional to. This method is used to measure high range in the process industry. The design of sensor is difficult [11].

The electrical property of water is analyzed based on change at different frequency range. If the excitation frequency increased the electrical constant increased. The basic electrical capacitance principle used to measure process parameter. The pressure is one of the factor to change of capacitance so they where consider the pressure range. Whose property of medium to analyze is placed in parallel plate of c If the frequency level is increased the conductance gets reduced as well permittivity gets reduced. [15].

## 2. PRINCIPLE OF ECT SYSTEM

The principle of the ECT system is that, when the dielectric medium between two conducting plate's changes, the capacitance between the

plates also changes. The ECT sensor usually consists of conducting plates connected in parallel or series. They can be considered as electrodes. These electrodes can be placed around the vessel (external connection) or it can be placed internally. The relationship existing between the capacitance and permittivity is governed by the equation:

$$C = Q/V = \epsilon_0 \cdot \epsilon_r \cdot (A/d) \quad (1.1)$$

Where,  $\epsilon_0$  is permittivity of vacuum.

$\epsilon_r$  is relative permittivity of medium inside the sensor.

d is distance between the plates.

A is area of the plates.

The value of  $\epsilon_0 \cdot \epsilon_r$  is the global average of fluid dielectric property over the entire sensing volume of the sensor known as permittivity. If the area of the plate and the distance between them are known, then we can measure the value of dielectric constant. This equation implies that capacitance is directly proportional to permittivity of the medium. A capacitor consists basically of two conductive plates separated by an insulating layer called the dielectric. The purpose of a capacitor is to store energy in the form of an electrical charge, Q on its plates. The ac capacitor will alternately charge and discharge at a rate determined by the frequency of the supply. This instantaneous change in voltage across the capacitor is opposed by the fact that it takes a certain amount of time to deposit (or release) this charge onto the plates and is given by  $V = Q/C$ . So we now know that capacitors oppose changes in voltage with the flow of electrons onto the plates of the capacitor being directly proportional to the rate of voltage change across its plates as the capacitor charges and discharges. Unlike a resistor where the opposition to current flow is its actual resistance, the opposition to current flow in a capacitor is called Reactance. Capacitive reactance of a capacitor decreases as the frequency across its plates increases. Therefore, capacitive reactance is inversely proportional to frequency.

Capacitive reactance opposes current flow, but the electrostatic charge on the plates (its AC capacitance value) remains constant. This means it becomes easier for the capacitor to fully absorb the change in charge on its plates during each half cycle. Also as the frequency increases, the current flowing into the capacitor increases in value because the rate of voltage change across its plates increases.

### 3. DESIGNING OF SENSOR

The modeling of capacitor sensor depends on the cross section of the pipe used around which the sensors are wound. The sensor cross section can be of any shape square or circular. The sensor contains an array of electrodes external to the pipe/vessel or internal to the pipe/vessel. The following figures represent how capacitors sensor is built surrounding the pipes in circular and square model. The sensor contains eight capacitance electrodes. These electrodes are made of copper whose permittivity is 6. Due to low frictional loss, circular pipes are preferred. The pipe is made of PVC pipe (Poly Vinyl Chloride) whose permittivity is 3.4. In order to improve the accuracy of imaging there are two ways: 1. Increase the number of measurement by increasing number of electrodes. 2. Improve the image reconstruction algorithm in order to obtain more information from captured data. If we increase the number of electrodes, the signal level will be low. As a result, the sensitivity in the center of imaging area will decrease.

#### 3.1 Sensor arrangement in Tank

The number of sensor electrodes that can be used depends on the range of values of inter-electrode capacitances.

##### 3.1.1 Sensor specifications

Total height of tank	=	56cm
Diameter of the tank	=	74cm
Length of the electrode	=	8cm
Height of the electrode	=	5cm




Fig.1 ECT Sensor arrangement

The ECT sensor consist of eight electrode placed on tank with insulator. The insulator protects damage and irresponsive of electrode from high temperature.

### 4. PROPOSED METHOD

The proposed system consist of tank with insulating layer and electrodes ,heater, AC execution source, embedded system, signal conditioning circuit ,DC voltmeter. Initially maintain the reference temperature to be measured at which the AC signal applied to the electrodes. If the electrode output is get based on permittivity of medium. They were used ECT system to measure the temperature at various range. Permittivity of medium is increase proportional to temperature and conductance decreases. Embedded system is switched electrode logic to measure various combinations then the signal conditioning circuit is used to convert current to voltage finally displayed in dc voltmeter. Here we measured temperature range is (25-70) Degree Celsius the equivalent voltage is varied (3-5)  $\mu\text{v}$  based on permittivity of medium. Temperature measurement is most important in parameter in thermal power plant to produce quality output and most of the temperature measurements are contact type in which measurement is difficult to measure at high temperature range. Thermocouple is very important role in temperature measurement and the one time used thermocouple to measure the high temperature measurement for accuracy then pyrometer is used high temperature with non-contact type. Low temperature measurement

thermostat RTD and LM35 sensor used .In our project both high and low temperature measurement used ECT sensor only and analysis the medium through image re construction. it is simple measurement and control technique used to remote control process.




Fig.2 Block Diagram of temperature measurement

### 5. EXPERIMENTAL WORK

The 8 sensor electrodes are wound around the tank. The electrodes are rectangular in shape, equally spaced and have same area of cross section. Sensors used in ECT system are designed according to the cross section of the vessel and positioning of electrode. The tank of the proposed system was non- conducting or conducting material and the sensors wound around the tank which is conducting plates. Thus the sensors are non- intrusive and easy to design. The functional base of an ECT system lies in the fact that the changes of measured capacitance values will depend on the material distribution inside the tank. Thus the system is not subjected to extreme and high pressure inside the tube. Also, the number of electrodes in a system is inversely proportional to image acquisition rate and overall resolution. However change in capacitance is directly proportional to change in permittivity inside the tank. The permittivity of medium is varied based on of medium. The proposed system is shown in figure shown below.




Fig.3 Hardware implementation

### 5.1 Result and Discussion

AC ECT system output is clearly shown in the tables. The permittivity of the air medium is low compared with water (liquid) and soil (solid) medium. Based on permittivity, medium yields the system output voltages. This assessment is used to measure the level of the medium inside the vessel. For an 8-electrode system, there are 28 unique voltage values obtained from different combination of electrode pair which acts as source and detector. The voltage values have a linear relation with capacitance in the presented system. Thus the correspondingly there will be 28 unique capacitance values, that are mapped onto a square grid containing 1024 pixels (32 pixel x32pixel) to graphically display the permittivity distribution inside the tank. The result shows water and soil permittivity inversely proportional to the temperature. The output voltages give equivalent temperature after calibration of ECT system.

Table 1: voltage values at 20 degree (c)

4.64	4.51	4.45	4.32	4.4	4.47	4.62
0	4.52	4.47	4.36	4.31	4.35	4.44
0	0	4.58	4.51	4.36	4.29	4.41
0	0	0	4.33	4.3	4.22	4.18
0	0	0	0	4.43	4.36	4.32
0	0	0	0	0	4	4.28
0	0	0	0	0	0	4.46






Fig.4 voltage values at 20 degree (c)




Fig.5 Voltages vs Temperature in Water and soil medium

## 6. CONCLUSION

This paper has presented the AC ECT system using analyzing and measurement. In the proposed method, senses dielectric medium soil and water based on permittivity distribution. Every medium has own permittivity, which measures of the medium. The AC method helps to increase system gain and system speed; it limits effect of stray capacitance. The experimental results show that this system detected the solid and liquid of the temperature inversely to the permittivity of the medium. This set of output value could be used to reconstruct the original image of the medium inside the tank. This system provides radiation-free, less expensive, non-intrusive and non-contact temperature measurement. We can look forward

high accuracy measurement investigate for various medium.

## References

- [1]. Zhaoyan Fan ,Robert X. Gao, A New Sensing Method for Electrical Capacitance Tomography, IEEE, 2010
- [2]. Nicolas Ticaud, Sophie Kohler, Pierre Jarrige, Lionel Duvillare”Specific Absorption Rate Assessment Using Simultaneous Electric Field and Measurements”IEEE antennas and wireless propagation letters, vol. 11, 2012.
- [3]. Simone Dalola, Vittorio Ferrar,Michele Guizzetti, Daniele Marioli”Autonomous Sensor System with Power Harvesting for Telemetric Measurements of Pipes”IEEE transactions on instrumentation and measurement, vol. 58, no. 5, may 2009.
- [4]. Hyungdae Bae, David Yun, Haijun Liu, Douglas A. Olson, and Miao Yu”Hybrid Miniature Fabry–Perot Sensor with Dual Optical Cavities for Simultaneous Pressure and Measurements”journal of light wave technology, vol. 32, no. 8, April 15, 2014.
- [5]. Akira Kimoto, Katsunori Shida”Imaging of -Change Distribution in the Brain Phantom by Means of Capacitance Measurement”IEEE transactions on instrumentation and measurement, vol. 49, no. 3, June 2000.
- [6]. Bruno Andò, Salvatore Baglio, Nicolò Savalli, and Carlo Trigona”Cascaded “Triple-Bent-Beam” MEMS Sensor forContactless Measurements inNonaccessible Environments”IEEE transactions on instrumentation and measurement, vol. 60, no. 4, April 2011.
- [7]. H.S. Tapp, A.J. Peyton, E.K. Kemsley, R.H. Wilson,” Chemical engineering applications of electrical process tomography” Sensors and Actuators B 92 (2003) 17–24.

- [8]. L. F. M. Moura, E. Cenedese and A. C. Azevedo Filho “Numerical study of a capacitive tomography system for multiphase flow” thermal engineering vol.8.no.02.December.
- [9]. QussaiMarashdeh, WarsitoWarsito, Liang-Shih Fan, and Fernando L. Teixeira” A Multimodal Tomography System Based on ECT Sensors,” IEEE sensors journal, vol. 7, no. 3 march 2007.
- [10]. W. Q. Yang, A. L. Stott, and J. C. Gamio,” Analysis of the Effect of Stray Capacitance on an AC-Based Capacitance Tomography Transducer “IEEE transactions on instrumentation and measurement, vol. 52, no. 5, October 2003.
- [11]. Wuliungyin, AnthonyJ.peyton, Grzgorzzyisko, and Richalddenno”IEEE transactions on Instrumentation and measurement, vol.57no, 11, November 2008.
- [12]. Zhaoyan Fan, Robert X. GAO,” Enhancement of Measurement Efficiency for Electrical Capacitance Tomography” IEEE transactions on instrumentation and measurement, vol. 60, no. 5, may 2011
- [13]. ZHayanfan, Robert x-gao”A Frequency selection scheme for increased imaging speed in ECT”: IEEE 2012.
- [14]. Li Lanying, Gao Ming and Chen Deyun, “A Novel Multiple Electrodes Excitation Method For Electrical Capacitance tomography System.” PP1167-1171, August 22-24, 2011
- [15]. Ji ying, Design and implementation of hardware system for electrical Capacitance Tomography first international workshop on education technology and computer science, pp(588-594) 2009.

**Manikandan. K** has completed B.E Degree in Electronics and communication Engineering (ECE) from Regional Engineering College, Tiruchirapalli in 2002, M.E Communication Systems (CSE) degree from Anna University,



Tiruchirapalli in 2009. He is doing Ph.D. in Anna University, Chennai, Tamilnadu. He is now working as Assistant professor in the department of Electronics and Instrumentation Engineering, J.J. College of Engineering and Technology, Tiruchirappali, Tamilnadu. His research interest Design of Electrical Capacitance Tomography with Embedded System for Industrial Application. He is a life member of IETE and ISTE.

**Sathiyamoorthy. S** has completed B.E Degree in Electrical and Electronics Engineering (EEE) in Regional Engineering College, Tiruchirapalli in 1995, M.E Control and Instrumentation (CIE) degree in Guindy Engineering College, Anna University, Chennai in 2002.

He has received his Doctorate degree from National Institute of Technology, Tiruchirapalli in 2008. He is now working as a Principal of J.J. College of Engineering and Technology, Tiruchirappali, Tamilnadu. His research areas include Process Control and Industry Automation, Multiphase flow identification using Electrical Capacitance Tomography. He is an active member of IEEE and a life member of ISTE and the International Society of Automation (ISA).

